

Etienne Cavalier*, Harjit P. Bhattoa, Annemieke C. Heijboer, Giovanni Lombardi, Vincent Delatour, Katleen Van Uytanghe, Konstantinos Makris, Niklas Rye Jørgensen, Markus Herrmann, Richard Pikner, Pierre Lukas, Samuel D. Vasikaran and Callum G. Fraser, on behalf of the IFCC-IOF Committee on Bone Metabolism

Assessment of measurement uncertainty of immunoassays and LC-MS/MS methods for serum 25-hydroxyvitamin D

<https://doi.org/10.1515/cclm-2025-0554>

Received May 8, 2025; accepted July 10, 2025;

published online August 12, 2025

Abstract

Objectives: Accurate 25-hydroxyvitamin D (25-(OH)D) assays are essential for defining vitamin D status and ensuring appropriate clinical decisions. Standardization efforts, including the Vitamin D Standardization Program (VDSP), aim to minimize assay variability. This study evaluates the measurement uncertainty (MU) of various 25-(OH)D assays and their ability to detect physiologically relevant changes over time.

Methods: Seventeen pooled and eight single-donor serum samples were analyzed using two LC-MS/MS methods and 13 immunoassays, each applied in two independent laboratories. Imprecision, bias, and MU were assessed relative to the University of Ghent's reference measurement procedure (RMP). Results were compared against analytical performance specifications (APS) from VDSP, JCTLM-TF-RMSI, and IFCC C-BM based on physiological 25-(OH)D variation. A

graphical approach was introduced to visualize MU in relation to clinical relevance.

Results: LC-MS/MS methods consistently met all APS criteria. Several immunoassays also achieved acceptable MU, although significant bias or inter-laboratory variability was observed for some of them. Slightly more than half of the assays met the desirable Joint Committee for Traceability in Laboratory Medicine Task Force on Reference Measurement System Implementation (JCTLM TF-RMSI) MU threshold ($\leq 10\%$), while four exceeded the minimum acceptable limit ($\leq 15\%$). The IFCC C-BM physiological approach identified a similar subset of assays. The graphical representation effectively illustrated method reliability across the tested concentration range.

Conclusions: Measurement uncertainty remains a major challenge for 25-(OH)D assays. The integration of MU-based APS and graphical visualization provides a comprehensive framework for evaluating assay performance. These findings highlight the importance of selecting assays capable of reliably detecting clinically meaningful changes in vitamin D status.

***Corresponding author: Etienne Cavalier**, Department of Clinical Chemistry, University of Liège, CIRM, CHU de Liège, Liège, Belgium, E-mail: etienne.cavalier@chuliege.be. <https://orcid.org/0000-0003-0947-2226>

Harjit P. Bhattoa, Department of Laboratory Medicine, Faculty of Medicine, University of Debrecen, Debrecen, Hungary. <https://orcid.org/0000-0002-4909-0065>

Annemieke C. Heijboer, Department of Laboratory Medicine, Endocrine Laboratory, Amsterdam UMC, Location Vrije Universiteit Amsterdam and Location University of Amsterdam, Amsterdam, The Netherlands

Giovanni Lombardi, Laboratory of Experimental Biochemistry, IRCCS Ospedale Galeazzi-Sant'Ambrogio, Milan, Italy. <https://orcid.org/0000-0002-8365-985X>

Vincent Delatour, Laboratoire National de Métrologie et D'Essais (LNE), Paris, France

Katleen Van Uytanghe, Department of Bioanalysis, Faculty of Pharmaceutical Sciences, Ref4U-Laboratory of Toxicology, Ghent University, Ghent, Belgium

Konstantinos Makris, Laboratory for the Research of Musculoskeletal System "Th. Garofalidis", School of Medicine, National and Kapodistrian,

University of Athens, Athens, Greece. <https://orcid.org/0000-0002-7896-9028>

Niklas Rye Jørgensen, Department of Clinical Biochemistry, Copenhagen University Hospital-Rigshospitalet, Copenhagen, Denmark; and Department of Clinical Medicine, Faculty of Health and Medical Sciences, University of Copenhagen, Copenhagen, Denmark

Markus Herrmann, Clinical Institute of Medical and Chemical Laboratory Diagnostics, Medical University of Graz, Graz, Austria

Richard Pikner, Department of Clinical Biochemistry and Bone metabolism, Klatovska Hospital, Klatovy, Czech Republic; Department of Clinical Biochemistry and Haematology, Faculty of Medicine Pilsen, Charles University Prague, Pilsen, Czech Republic; and Faculty of Health Care Studies, University of West Bohemia, Pilsen, Czech Republic

Pierre Lukas, Department of Clinical Chemistry, University of Liège, CIRM, CHU de Liège, Liège, Belgium

Samuel D. Vasikaran, PathWest Laboratory Medicine, Fiona Stanley Hospital, Murdoch, WA, Australia

Callum G. Fraser, Population Health and Genomics, Ninewells Hospital and Medical School, University of Dundee, Dundee, UK. <https://orcid.org/0000-0002-1333-7994>

Keywords: 25-hydroxyvitamin D; measurement uncertainty; standardization; LC-MS/MS; immunoassays; analytical performance specifications

Introduction

Standardization of total 25-hydroxyvitamin D assays is essential for ensuring consistent clinical and epidemiological interpretations [1]. Standardization efforts aim to minimize inter-assay variability, establish reliable concentration thresholds for defining vitamin D status, and provide a framework for vitamin D supplementation guidelines. The Vitamin D Standardization Program (VDSP) and its Vitamin D Standardization-Certification Program (VDSCP), led by the Centers for Disease Control and Prevention (CDC) and the International Federation of Clinical Chemistry and Laboratory Medicine (IFCC) [2], are key initiatives promoting standardization and method comparability across laboratories. These programs have supported the development of three Joint Committee for Traceability in Laboratory Medicine (JCTLM) recognized reference measurement procedures (RMPs), which are maintained at the National Institute of Standards and Technology (NIST) [3], Ghent University [4], and the CDC [5].

To ensure meaningful standardization, analytical performance specifications (APS) must be established. The APS historically applied in vitamin D measurement standardization are those proposed by Stöckl et al. [6], as endorsed by the clinical laboratory community and agreed upon at the 1999 Stockholm Consensus Conference on Analytical Goals in Laboratory Medicine [7]. The VDSP has incorporated these APS into its certification process, whereby a 25-(OH)D method is considered “certified” if its calibration bias remains within $\pm 5\%$ of the CDC RMP and its imprecision, expressed as the coefficient of variation (CV), is $< 10\%$ [8]. The VDSCP also evaluates individual sample pass rates, i.e., the percentage of samples meeting bias criteria. Recent evaluations have shown that LC-MS/MS methods achieve higher pass rates (40–92 %, mean 65 %) compared to immunoassays (10–45 %, mean 27 %), highlighting the substantial variability between methods [9].

While standardization efforts focus primarily on bias and imprecision, measurement uncertainty (MU) provides a complementary approach for evaluating assay performance. MU integrates both systematic and random errors, offering a more comprehensive assessment of an assay’s reliability. In this manuscript, MU refers to the expanded measurement uncertainty (U), calculated from the combined standard uncertainty (u_c), which itself incorporates both standard uncertainty due to imprecision (u_{IMP}) and bias (u_{BIAS}), as defined by ISO 20914. The JCTLM Task Force on Reference

Measurement System Implementation (JCTLM TF-RMSI) has proposed APS specifically targeting combined MU, defining a desirable threshold of $\leq 10\%$ and a minimum acceptable threshold of $\leq 15\%$ for clinical samples [10]. Additionally, the IFCC–IOF Committee on Bone Metabolism (IOF-IFCC C-BM) has also proposed a complementary approach based on MU to derive APS for 25-(OH)D based on the 2014 Milan Strategic Conference “Model 1” [11]. The APS we proposed are based on the physiological variation of 25-(OH)D concentrations over time. Specifically, data from the European Biological Variation Study (EuBIVAS) [12] demonstrated that, in healthy individuals, 25-(OH)D concentrations naturally increased by 31.6 % over a 10-week period (April–June) [13]. This increase is primarily driven by seasonal effects, where progressive sun exposure during spring and summer enhances cutaneous vitamin D synthesis, leading to a biologically expected rise in circulating 25(OH)D levels. To evaluate whether analytical methods can reliably detect such physiologically meaningful changes, an assay should exhibit MU that allows for statistically significant differentiation of this increase. The minimum required MU for this purpose can be determined using the following formula:

$$\text{Change (\%)} = \sqrt{2} \times \text{MU (\%)} \times Z$$

where Z represents the Z-score, corresponding to different levels of statistical confidence. Since this variation represents an increase, the statistical approach must be one-sided. According to Fraser [14], the probability of detecting a significant rise can be classified as.

- Likely detection ($p > 0.80$) $\rightarrow Z = 0.84$
- More than likely detection ($p > 0.90$) $\rightarrow Z = 1.28$
- Very likely detection ($p > 0.95$) $\rightarrow Z = 1.65$
- Virtually certain detection ($p > 0.99$) $\rightarrow Z = 2.33$

Thus, an assay capable of detecting a 31.6 % increase in 25-(OH)D over 10 weeks with these probabilities must have an MU of $\leq 26.5\%$ (likely detection), $\leq 17.4\%$ (more than likely detection), $\leq 13.6\%$ (very likely detection) and $\leq 9.6\%$ (virtually certain detection), respectively.

In this multicentric study, we used the University of Ghent (UGhent) RMP to evaluate the performance of two distinct LC-MS/MS methods and 13 immunoassays. Performance was compared against the APS defined by the VDSP, JCTLM TF-RMSI and IOF-IFCC C-BM approach based on physiological variation of 25-(OH)D over 10 weeks. Finally, we introduced a novel graphical approach based on MU, providing an intuitive visualization of method performance to complement traditional APS evaluations.

Materials and methods

Preparation of samples

A collection of 17 pooled serum samples was prepared by pooling at least 10 remnant fresh samples whose concentration had been previously estimated with the DiaSorin Liaison method, routinely used at the CHU de Liège. The selection of the samples to create the collection was designed to cover a wide measurement range, spanning from the LOQ of most methods to approximately 100 µg/L. Based on their concentration, selected samples were pooled into seventeen 100 mL tubes (Greiner BioOne, REF 227261, Frickenhausen, Germany), which were stored at -80°C . The collection of samples ceased once each vial contained approximately 50 mL of serum. It took approximately three months to collect the samples to prepare the pools. Once the desired volumes were achieved, pools were thawed at $+4^{\circ}\text{C}$, vortexed for 15 s and homogenized for 30 minutes on an orbital shaker at room temperature. They were then centrifuged for 10 min at 4,000 g (39.240 m/s^2). After centrifugation, aliquots of 500 µL were prepared in 2 mL screw-capped tubes (Eppendorf Protein LoBind, REF 0030108132, Oldenburg in Holsten, Germany) for the samples dedicated to LC-MS/MS analysis and Sarstedt micro tubes (REF 72.694.007, Nümbrecht, Germany) for the immunoassays. These aliquots were stored at -80°C until shipment to participating laboratories and the University of Ghent on dry ice or measurement in the CHU de Liège laboratory.

A second set of samples was prepared by using eight single donor samples which were purchased from SLR Research Corporation (Carlsbad, CA). Each donation contained a portion of 250 mL single donations consisting of separated serum which were selected based on the nominal 25-(OH)D concentration provided by the manufacturer. The donations were shipped on dry ice and the aliquots were prepared following the same process as described above.

Two aliquots of 500 µL of each sample were shipped to each participating laboratory to allow a measurement in duplicates on two consecutive days ($n=4$ determinations per sample).

The University of Ghent measured the samples in duplicates with the RMP for their 25-(OH)D3 concentration reporting results in ng/g. 25-(OH)D2 was not measured because of the limited sample volume available.

Examination methods and participating laboratories

Two LC-MS/MS methods, developed by the CHU de Liège (BE) and the Amsterdam UMC (NL), were used in this study. These

methods have been described elsewhere [15, 16]. Of note, the method of the CHU de Liège is recognized as meeting the standardization criteria for traceability to the JCTLM RMPs by the CDC VDSCP [9]. In addition, the LC-MS/MS method from CHU de Liège was also used to quantify 24,25-dihydroxyvitamin D and three-epi-25-hydroxyvitamin D₃ in the samples. These measurements were not part of the primary objective of the study and are therefore not included in the main analysis. The results are provided in Supplementary Table 1 for information.

Samples were also analyzed using various immunoassays across multiple clinical laboratories. The full list of the methods and their characteristics is detailed in Table 1. Most of the measurements were conducted by two different laboratories for each evaluated method, and manual methods were performed by two different experienced technicians when possible.

The examinations were performed using routine laboratory operating procedures under a total quality management system. Commercial immunoassays were operated according to the manufacturers' instructions, and all laboratories followed internal quality control procedures to ensure result validity. Results were reported in µg/L to one significant figure after the decimal point.

Data analysis

For each method, the mean and standard deviation (SD) of the four examinations were calculated, and the coefficient of variation (CV, %) was derived. The standard uncertainty due to imprecision u_{IMP} was calculated as: $u_{\text{IMP}} = \frac{\text{SD}}{(\sqrt{n}-1)}$, where n corresponds to the number of examinations ($n=4$). The absolute bias was determined as the difference between the mean of the four replicate measurements and the UGhent RMP assigned value.

Absolute bias = Mean of 4 replicates

– UGhent RMP assigned concentration

The standard uncertainty due to bias (u_{BIAS}) was then calculated as:

$$u_{\text{BIAS}} = \frac{\text{Absolute bias}}{\sqrt{(n-1)}}$$

The combined standard uncertainty (u_c) was then computed as.

$$u_c = \sqrt{(u_{\text{IMP}}^2 + u_{\text{BIAS}}^2)}$$

And the expanded absolute measurement uncertainty (U) was calculated as $U = u_c \times 2$

Table 1: Assay manufacturers, testing laboratories, and instrumentation used for 25-(OH)D measurements.

Examination manufacturer (short name in bold) with full name, country of origin	Site	Examination model	Examination type	Reagent lot no	Calibrator lot no.
Abbott Diagnostics, Germany	CHU de Liège, Belgium	Architect 25-OH vitamin D; ALINITY serie i	CLIA/CMIA	43057UD00	40423UD00A-F (6 levels)
Abbott Diagnostics, Germany	Amsterdam UMC, Endocrine Laboratory, The Netherlands	Architect 25-OH vitamin D; ALINITY serie i	CLIA/CMIA	45580UD00	45248UD00 (6 levels)
Beckman-Coulter , USA	CHU de Liège, Belgium	Access 25(OH) vitamin D total; Access-2	CLIA	234057	234224
Beckman-Coulter , USA	Dienst laboratoriumgeneeskunde AZ Delta, Roeselare, Belgium	Access 25(OH) vitamin D total; Access 2	CLIA	234057	234427
DiaSorin , Italy	CHU de Liège, Belgium	Liaison 25 OH vitamin D total; Liaison XL	CLIA	136088-4972	136088
DiaSorin , Italy	University of Debrecen, Department of Laboratory Medicine, Hungary	Liaison 25 OH vitamin D total; Liaison XL	CLIA	136263	136263
Diazyme , USA	CHU de Liège, Belgium	EZ vitamin D assay	PETIA	VL006210-10-01	VL502210-0,501
Fujirebio , Japan	CHU de Liège, Belgium	Cobas 6,000 c501	CLIA	8XX3031	8YX3052E
Fujirebio , Japan	Laboratoire Hospitalier Universitaire de Bruxelles, Belgium	25-OH vitamin D; Lumipulse G1200	CLIA	8XX3082	8YX3052E
Immunodiagnostic systems (IDS), UK	CHU de Liège, Belgium	25-OH vitamin D; Lumipulse G1200	CLIA	104261-0	104217-0
Immunodiagnostic systems (IDS), UK	Rigshospitalet Adeling for Klinisk Biokemi, Glostrup, Denmark	IDS-iSYS 25-hydroxy vitamin D	CLIA	103840-0	104217-0
Siemens Healthineers , USA	Groupe santé CHC-Laboratoire d'Analyses Médicales, Liège, Belgium	IDS-iSYS 25-hydroxy vitamin D	CLIA	25940190	C3490
Siemens Healthineers , USA	CHR Citadelle, Liège, Belgium	Vitamin D total (VitD); Atellica IM	CLIA	1851563	C3485
Roche Diagnostics, Germany	CHU de Liège, Belgium	Vitamin D total (VitD); Atellica IM	CLIA	64622701	63910401
Roche Diagnostics, Germany	University of Debrecen, Department of Laboratory medicine, Hungary	Vitamin D total II; Cobas e411	CLIA	64622702	639104
Snibe , China	CHU de Liège, Belgium	Vitamin D total II; Cobas e411	CLIA	262220211	262220211
Snibe , China	University of Debrecen, Department of Laboratory medicine, Hungary	25-OH vitamin D (CLIA) Maglumi x3	CLIA	262230111	262230111
Ortho Clinical Diagnostics , USA	Rigshospitalet Afdeling for Klinisk biokemi, Glostrup, Denmark	25-OH vitamin D (CLIA) Maglumi x3	CLIA	1830	175
Ortho Clinical Diagnostics , USA	CHU UCL Namur, Belgium	Vitros 5,600	CLIA	1730	0167
		VITROS 3600	CLIA		

Table 1: (continued)

Examination manufacturer (short name in bold) with full name, country of origin	Site	Examination model	Examination type	Reagent lot no	Calibrator lot no.
Immunodiagnostic Systems (IDS), UK	CHU de Liège, Belgium	25-Hydroxy vitamin D EIA (IDS); Thermo fischer multiskan FC with incubator	EIA	J52547	J52359-65 (7 levels)
Diasource , Belgium	CHU de Liège, Belgium	25OH vitamin D total ELISA; Thermo fischer multiskan FC with incubator	ELISA	2206–5986	2110-3612/02 2109-3402/02
Affimedix , USA	CHU de Liège, Belgium	25 OH quantitative vitamin D rapid test Cube reader	IC	22112182	22112182

This Table provides details on the immunoassays included in the study, including the manufacturer, country of origin, participating laboratory, examination model, examination type, reagent lot number, and calibrator lot number. Assays were conducted in multiple laboratories across different institutions. CLIA, chemiluminescent immunoassay; CMA, chemiluminescent microparticle immunoassay; PETIA, particle-enhanced turbidimetric immunoassay; EIA, enzyme immunoassay; ELISA, enzyme-linked immunosorbent assay; IC, immunochromatographic assay.

Finally, in order to express the uncertainty in relative terms, the expanded relative measurement uncertainty (U_{rel}), referred to here as CREMU (Combined Relative Expanded Measurement Uncertainty), was derived by normalizing the expanded absolute uncertainty to the UGhent RMP assigned value:

$$CREMU = \frac{U}{\text{UGhent RMP assigned concentration}} \times 100$$

Results

According to the UGhent RMP, serum total 25-(OH)D concentrations in the 17 pooled and eight single-donor samples ranged from 3.4 µg/L to 73.7 µg/L, with a mean of 31.7 µg/L. One sample contained a significant amount of 25-(OH)D2 (~5 µg/L), as measured by LC-MS/MS methods, and was excluded since UGhent RMP did not provide 25-(OH)D2 measurements in this study. Another low-concentration sample was removed because its 25-(OH)D3 level was at or below the LOQ of several methods. Additionally, one sample from the single-donor set had concentrations below the LOQ of the Ortho system (<8 µg/L) and was excluded from the evaluation of this method only.

The mean imprecision (CV%) and mean bias (%) for serum total 25-(OH)D based on four individual replicate measurements, as well as VDSP performance criteria by assay, are presented in Table 2. Briefly, both LC-MS/MS methods and four automated immunoassays (Beckman-Coulter, DiaSorin, Fujirebio and IDS iSYS and ELISA – even if in one laboratory out of two for these three latter) passed the VDSP criteria for bias and imprecision. The individual sample pass rate, representing the percentage of determinations with bias between ± 5 %, ranged from 0 % (for Diazyme and Roche in one laboratory) to 87 % for the UMC Amsterdam LC-MS/MS. Table 2 also provides a detailed breakdown of bias, imprecision, and measurement uncertainty across participating laboratories. Significant inter-laboratory variability was observed for certain methods. With the Abbott assay, one laboratory exhibited a bias nine times higher than the other, suggesting considerable variability in calibration or operational conditions. Similarly, for DiaSorin, the CV in one lab was three times higher than in the other, indicating substantial differences in precision. Among the two laboratories using Ortho Diagnostics, one displayed extreme bias above 60 %, which remained uncontrolled even after recalibration and reruns with new samples. In ELISA-based methods, operator variability was evident despite all technicians being experienced and using

Table 2: Bias, imprecision, and measurement uncertainty for serum 25-(OH)D assays across participating Laboratories.

Method	Laboratory	Mean CV, %	Mean bias, %	Pass VDSP APS (bias < 5 % and imprecision <10 %)	Ample pass rate (% individual values with bias <±5 %)	Unexpanded MU, %	CREMU, %	Pass JCTLM TF-RMSI desirable APS for standard MU on clinical samples (≤10 %) [10]	Pass JCTLM TF-RMSI acceptable APS for standard MU on clinical samples (≤15 %) [10]	Capacity of detecting a 31.6 % change in 25-(OH) D over 10 weeks (IOF-IFCC C-BM approach)
LC-MS/MS										
CHU de Liège	A	3.1 %	3.5 %	YES	65.2 %	2.9 %	5.7 %	YES	YES	Virtually certain
Amsterdam UMC	B	3.7 %	2.9 %	YES	87.0 %	2.9 %	5.8 %	YES	YES	Virtually certain
Immunoassays										
Abbott	A	3.3 %	2.0 %	YES	30.4 %	5.1 %	10.2 %	YES	YES	Very likely
Abbott	B	3.5 %	18.6 %	NO	8.7 %	11.1 %	22.1 %	NO	YES	Likely
Beckman-Coulter	A	5.3 %	-2.9 %	YES	34.8 %	7.2 %	14.4 %	YES	YES	More than likely
Beckman-Coulter	C	5.1 %	-3.1 %	YES	30.4 %	6.9 %	13.7 %	YES	YES	More than likely
DiaSorin	A	3.2 %	-0.8 %	YES	43.5 %	5.0 %	10.0 %	YES	YES	Very likely
DiaSorin	D	9.9 %	4.6 %	YES	34.8 %	8.0 %	16.0 %	YES	YES	More than likely
Fujirebio	A	1.7 %	-5.6 %	NO	56.2 %	3.4 %	6.7 %	YES	YES	Virtually certain
Fujirebio	E	3.9 %	3.9 %	YES	47.8 %	4.0 %	8.0 %	YES	YES	Virtually certain
IDS	A	2.4 %	5.2 %	NO	52.2 %	4.2 %	8.4 %	YES	YES	Virtually certain
IDS	F	4.9 %	1.8 %	YES	65.2 %	4.3 %	8.6 %	YES	YES	Virtually certain
Roche	A	5.3 %	11.5 %	NO	21.7 %	10.4 %	20.7 %	NO	YES	Likely
Roche	D	7.0 %	17.1 %	NO	0.0 %	12.6 %	25.2 %	NO	YES	Likely
Siemens	G	4.1 %	28.9 %	NO	13.0 %	17.9 %	35.7 %	NO	NO	Not likely
Siemens	H	6.3 %	20.1 %	NO	21.7 %	14.8 %	29.5 %	NO	YES	Not likely
Snibe	A	1.4 %	11.3 %	NO	17.4 %	6.7 %	13.4 %	YES	YES	Very likely
Snibe	D	4.6 %	5.6 %	NO	34.8 %	6.6 %	13.1 %	YES	YES	Very likely
Ortho	F	7.6 %	8.6 %	NO	13.0 %	10.6 %	21.2 %	NO	YES	Likely
Ortho	H	6.7 %	64.2 %	NO	4.5 %	38.0 %	75.9 %	NO	NO	Not likely
IDS ELISA	A ₁	8.7 %	11.1 %	NO	30.4 %	10.4 %	20.8 %	NO	YES	Likely
IDS ELISA	A ₂	7.1 %	4.4 %	YES	26.1 %	9.8 %	19.6 %	YES	YES	Likely
DiaSource ELISA	A ₁	16.0 %	12.9 %	NO	30.4 %	15.9 %	31.8 %	NO	NO	Not likely
DiaSource ELISA	A ₂	5.6 %	17.3 %	NO	13.0 %	12.1 %	24.2 %	NO	YES	Likely

Table 2: (continued)

Method	Laboratory	Mean CV, %	Mean bias, %	Pass VDSP APS (bias < 5 %) and imprecision < 10 %	Ample pass rate (% individual values with bias < ± 5 %)	Unexpanded MU, %	CREMU, %	Pass JCTLM TF-RMSI desirable APS for standard MU on clinical samples (≤ 10 %) [10]	Pass JCTLM TF-RMSI minimum acceptable APS for standard MU on clinical samples (≤ 15 %) [10]	Capacity of detecting a 31.6 % change in 25-(OH) D over 10 weeks (IOF-IFCC C-BM approach)
Diazyme	A	3.0 %	−20.3 %	NO	0.0 %	12.4 %	24.8 %	NO	YES	Likely
Affimedix	A ₁	7.7 %	35.1 %	NO	13.0 %	25.4 %	50.7 %	NO	NO	Not likely

This Table presents the mean coefficient of variation (CV, %), mean bias (%), percentage of individual values with bias < ± 5 % (sample pass rate), unexpanded measurement uncertainty (unexpanded MU, %), and combined relative expanded measurement uncertainty (CREMU, %) for each assay. Performance was assessed according to the Vitamin D Standardization Program (VDSP) criteria (bias within ± 5 % and imprecision < 10 %) and MU, thresholds based on established analytical performance specifications (APS). The study evaluated 17 freshly prepared pools of serum samples and eight serum donations from healthy volunteers. The capacity of each method to detect a 31.6 % increase in 25-(OH)D concentrations over 10 weeks is categorized as Not likely (p < 0.80), Likely (p > 0.80), More than likely (p > 0.90), Very likely (p > 0.95), Virtually certain (p > 0.99). A, CHU de Liège; A₁ and A₂, experienced technicians from the CHU de Liège; B, Amsterdam UMC; C, AZ Delta; D, University of Debrecen; E, Laboratoire Hospitalier Universitaire de Bruxelles; F, Rigshospitalet; G, CHR Citadelle; Groupe CHC; H2CCCHU UCL Namur. APS, analytical performance specifications; VDSP, vitamin D standardization program; MU, measurement uncertainty; CREMU, combined relative expanded measurement uncertainty; CV (%), coefficient of variation; bias (%), systematic error deviation from the reference measurement procedure; LC-MS/MS, liquid chromatography-tandem mass spectrometry; ELISA, enzyme-linked immunosorbent assay.

identical instruments under controlled environmental conditions.

To further assess assay performance, measurement uncertainty of the methods was also evaluated against the different APS frameworks described above. The two LC-MS/MS and the methods from Beckman, DiaSorin, Fujirebio, IDS iSYS and Snibe met the JCTLM TF-RMSI desirable APS for standard MU on clinical samples (≤ 10 %) in all the laboratories whereas the methods from Abbott and the IDS ELISA were successful in one out of two centres; All the other methods failed to meet these APS. The “minimum acceptable” APS, set at (≤ 15 %) was met by all methods, except four of them: Siemens, Ortho and Diasource (in one out of two participating laboratories) and Affimedix.

Figure 1 presents, for all methods combined, a set of correlation graph where the x-axis represents $x(UGhentRMP)$ and the y-axis represents the mean of the 4 replicates for each sample included in the study. Each data point is displayed with an error bar corresponding to the calculated unexpanded MU (%). V-shape tolerance limits were applied to each graph at $x(UGhent) \pm 9.6 \%$, $x(UGhent) \pm 13.6 \%$, $x(UGhent) \pm 17.4 \%$ and $x(UGhent) \pm 26.5 \%$. These limits represent the MU required for an analytical method to detect a 31.6 % physiological change in 25-(OH)D concentrations after a 10-weeks period with a probability of 99, 95, 90 and 80 %, respectively. The V-shaped zones provide a graphical representation of each method's ability to track physiologically relevant changes. An assay is considered more clinically useful if its data points and MU bars fall within the narrower zones (e.g., within the $\pm 9.6 \%$ or $\pm 13.6 \%$ limits). LC-MS/MS methods consistently remained within the most stringent V-shape boundaries, confirming their ability to reliably track changes in vitamin D concentrations. In contrast, several immunoassays exceeded the broader V-shaped limits, indicating that their MU might compromise their ability to detect meaningful fluctuations in 25-(OH)D levels over time. Detailed individual plots for each method are provided in Supplementary Figures S1–S13.

An alternative graphical representation of the results is also shown in Figure 1. In this approach, the CREMU (%) is plotted on the y-axis against the x-axis, which represents UGhent RMP. The mean CREMU (%) is displayed for each method, along with horizontal reference lines corresponding to the previously mentioned limits. This graph effectively illustrates how the methods behave across different concentrations and allows for an immediate comparison between pooled samples and single donations, as well as differences between methods.

For instance, Abbott shows a significant increase in CREMU at higher 25-(OH)D concentrations in one out of the two laboratories, likely due to high cross-reactivity with

other vitamin D metabolites, such as 24,25-(OH)₂ vitamin D. In the case of Ortho, the discrepancy between the two laboratories is quite pronounced, possibly due to an incorrect calibrator value assignment in one of the lots. Siemens exhibits a notably high CREMU in the low concentration range, particularly for single donations, which may be attributed to a matrix effect – though the pools with similar low concentrations also show elevated CREMU. Lastly, the figure highlights the excellent performance of LC-MS/MS methods, alongside certain immunoassays such as Fujirebio and IDS.

Discussion

In this study, we have used eight single donations of serum and 17 pools made from remnant samples spanning the measurement range to compare the performances of thirteen immunoassays and two distinct LC-MS/MS methods with the RMP from the University of Ghent. The samples

have been measured in duplicate on two consecutive days in two different laboratories and we calculated the imprecision (CV%), the bias and the measurement uncertainty for each data point obtained with each method.

The 25-(OH)D imprecision and bias of most of these methods had previously been estimated by Wise et al. according to the VDSP APS (mean bias <|5 %| and mean CV<10 %) in 50 single-donor samples containing 25-(OH)D₂ and 25-(OH)D₃ or 25-(OH)D₃ alone [17]. The results showed that three of the 12 methods they tested failed to pass, mainly due to a mean bias that exceeded the allowable limit of |5 %|. In our study, both LC-MS/MS, Beckman and DiaSorin were able to pass the VDSP criteria whereas this was the case for one lab (or technician) out of two for Fujirebio, Abbott, IDS iSYS and IDS ELISA. All the other methods failed. The reason for failing to pass the VDSP was the bias, which was higher than ± 5 %. Of note, one of the labs failed with Fujirebio while presenting a CV of 1.7 % and a bias of -5.6 % whereas a lab using DiaSorin succeeded with a CV of 9.9 % and a bias of 4.6 %.




Figure 1: (A) Combined relative expanded measurement uncertainty (CREMU) across all immunoassays and LC-MS/MS methods. The CREMU (%) is plotted on the y-axis against the UGhent reference measurement procedure (RMP)-assigned 25(OH)D concentrations on the x-axis. Each method's mean CREMU is shown, along with horizontal reference lines corresponding to the predefined analytical performance specifications (APS). This panel illustrates how methods perform across the concentration range and allows direct comparison between pooled samples (red dots) and single donations (black dots). (B) Combined correlation graphs with measurement uncertainty (MU) and V-shaped acceptance zones for all methods. Each data point represents the mean of four replicates per sample, with the x-axis showing the UGhent RMP-assigned value and the y-axis showing the method-specific mean. Vertical error bars indicate the calculated unexpanded MU (%). Superimposed V-shaped zones delineate thresholds of ± 9.6 %, ± 13.6 %, ± 17.4 %, and ± 26.5 % around the RMP value, reflecting the MU required to detect a 31.6 % physiological change in 25(OH)D concentrations over 10 weeks with probabilities of 99, 95, 90, and 80 %, respectively. Methods with data points and MU bars within narrower zones (e.g., ± 9.6 % or ± 13.6 %) are more likely to detect clinically meaningful changes reliably. LC-MS/MS methods consistently remain within the tightest zones, while some immunoassays exceed the broader limits, indicating potential limitations for longitudinal monitoring. Red dots represent pooled samples; black dots indicate individual single donations. Detailed method-specific plots are provided in Supplementary Figures S1–S13.

The sample pass rate, which is an interesting metrics proposed by the VDSCP to evaluate the percentage of individual samples that pass the VDSP APS, is also of interest. If the two LC-MS/MS methods and one lab using IDS iSYS presented a high sample pass rate (from 65.2 to 87.0 %), the percentage dropped to around 50 % for the other iSYS system and Fujirebio to 30 % for the tests who succeeded the VDSP (Beckman and DiaSorin) and 20 % or lower for the other methods.

The JCTLM TF-RMSI has recently proposed APS for vitamin D measurement, establishing desirable and minimum thresholds for measurement uncertainty at $\leq 10\%$ and $\leq 15\%$, respectively [10]. We applied these APS to evaluate the ability of different methods to meet these criteria. Our findings show that slightly more than half of the tested methods met the $\leq 10\%$ target, while only four methods failed to stay within the $\leq 15\%$ limit, suggesting that this threshold may be relatively permissive. When applying the IFCC-IOF C-BM approach-assessing a method's likelihood of detecting a physiological change based on its uncertainty [13] - we found a strong alignment with the TF-RMSI APS. Specifically, methods classified as virtually certain, very likely, or more than likely to detect this change corresponded to those meeting the $< 10\%$ APS, whereas those classified as only likely (or failing to detect the change) did not meet this APS.

Another key objective of this study was to assess the feasibility of using MU and novel graphical representations to visually evaluate assay performance. Traditional assessments of analytical performance often rely on separate evaluations of bias and imprecision and do not always provide an intuitive representation of whether an assay can reliably detect physiologically relevant changes in 25-(OH)D concentrations. By integrating MU into graphical representations, the approach we propose allows for a more holistic assessment of method performance, combining analytical errors into a single visualization. One of the major advantages of these graphs is that they provide an immediate and intuitive overview of an assay's reliability, making it easier to compare different methods at a glance. Instead of requiring a detailed examination of multiple performance metrics, laboratorians and clinicians can quickly assess whether an assay's MU is low enough to ensure the reliable detection of clinically relevant changes in vitamin D status. While numerical APS thresholds remain essential for standardization and certification, we believe that graphical representation enhances the interpretation of assay reliability by offering a more clinically meaningful perspective. Future standardization efforts should consider integrating both numerical MU assessments and intuitive visual tools to provide a more comprehensive evaluation of method performance.

Our study has however some weaknesses. It was conducted in only two laboratories per method, raising the question of whether these labs are fully representative of broader clinical and research settings. This is particularly relevant given that substantial inter-laboratory differences were observed for some immunoassays, despite their use under routine conditions. Also, while the two LC-MS/MS methods in this study showed comparable performance, in-house developed LC-MS/MS methods can exhibit significant variability in accuracy and precision across different laboratories. Therefore, caution should be exercised when generalizing these findings to other in-house LC-MS/MS methods that were not included in this study. Another limitation is that the commutability of the samples (i.e. the generated pools and the single donations) has not been established. Furthermore, most of the single donations had 25-(OH)D concentrations lower than $20\text{ }\mu\text{g/L}$, whereas of the majority of the pools exhibited concentrations higher than $20\text{ }\mu\text{g/L}$. Matrix effects are thus possible even though, visually, we did not observe overt discrepancies in behavior between the pools and the single donations. The nature of the anonymized samples, both single donations and pools (by definition), that were used in this study can also be questioned. The single donations came from apparently healthy subjects from the United-States, and no data were available on the ethnicity of the subjects. It is well known that vitamin D binding protein (VDBP) is highly polymorphic, and that allelic combinations differ in their affinities for vitamin D metabolites, circulating concentrations, and geographical and ethnic distribution [18]. The pools were made of remnant anonymized single serum samples coming from the routine practice of the CHU de Liege and it is very well possible that some of these samples were from patients suffering from CKD or hemodialyzed patients. It is well known that immunoassays for vitamin D can be affected by CKD [19, 20] which might explain differences observed in this study compared to external quality assessment schemes and similar studies using donations from healthy individuals only [21, 22]. Pregnancy, especially in the last trimester, can also affect VDBP concentrations [19], but the CHU de Liège does not have a maternity ward and there is highly improbable that a substantial number of samples coming from pregnant women were included in the pools.

Also, in this study, the estimation of MU was primarily based on a bottom-up approach, using within-laboratory imprecision (u_{Rw}) and bias relative to the RMP. Although ISO 20914 recommends a top-down approach incorporating all uncertainty sources along the traceability chain, including the uncertainty associated with reference materials (u_{ref}) and calibrator value assignment (u_{cal}), these data could not be easily retrieved. Moreover, no MU estimates are

provided by manufacturers in the kit inserts or validation documentation (a limitation we regret and believe should be addressed), making a full top-down estimation impractical. Since the bias component in our MU calculation is derived from comparison against the RMP, it implicitly covers part of the uncertainty linked to the calibration hierarchy. Therefore, for laboratories that do not design measurement procedures from scratch, a bottom-up estimation represents a pragmatic and acceptable approach to assess MU.

The study has also many strengths. The protocol we used is similar to the VDSCP and the study was performed in a total independence from the manufacturers, in a large number of experienced laboratories situated in various countries across Europe. Moreover, each method was run in at least two laboratories or two different technicians (except Affimedix). The results of this study should however not be over interpreted. As already mentioned, the excellent performance of the LC-MS/MS is not per definition representative for all LC-MS/MS methods and the performance of the immunoassays depicted in this study is only relevant for the lots of reagents and calibrators used.

Overall, this study highlights the persistent variability in 25-(OH)D assay performance despite ongoing standardization efforts. While LC-MS/MS methods have demonstrated superior accuracy and lower measurement uncertainty, certain immunoassays also met the established analytical performance specifications. The study underscores the importance of integrating measurement uncertainty into performance evaluations and introduces novel graphical representations to enhance the interpretation of assay reliability. However, limitations such as inter-laboratory variability, potential matrix effects, and the absence of commutability assessments should be considered when interpreting the findings. Future standardization initiatives should further explore the use of uncertainty-based metrics and visualization tools to improve method comparability and ensure clinically meaningful assessments of vitamin D status.

Acknowledgments: We would like to thank the colleagues who kindly accepted to measure the samples: Inge De Cuyper and Dieter Desmet (AZ Delta), Marie-Louise Schleck (Hôpital de la Citadelle), Lucas Dewalque (CHC Mont-Légia), Emilie Catry (CHU UCL Mont-Godinne), Fleur Wolf, and Frédéric Cotton (LHUB Bruxelles).

Research ethics: Not applicable, as the study used anonymized pooled serum samples, and no patient-identifiable information was collected or processed.

Informed consent: Not applicable.

Author contributions: All authors have accepted responsibility for the entire content of this manuscript and

approved its submission. Soft-enter replaced as Paramark, EC, HPB, ACH, GL, VD, KVV, KM, NRJ, MH, RP, PL, SDV, and CGF contributed to the design, data collection, analysis, and interpretation of the study. EC drafted the manuscript, and all authors critically revised it.

Use of Large Language Models, AI and Machine Learning Tools: None declared.

Conflict of interest: EC is a consultant for IDS, DiaSorin, Fujirebio, Snibe, and Roche Diagnostics. All other authors state no conflict of interest.

Research funding: None declared.

Data availability: The data supporting the findings of this study are available from the corresponding author upon reasonable request.

References

1. Sempos CT, Heijboer AC, Bikle DD, Bollerslev J, Bouillon R, Brannon PM, et al. Vitamin D assays and the definition of hypovitaminosis D: results from the first international conference on controversies in vitamin D. *Br J Clin Pharmacol* 2018;25.
2. Sempos CT, Vesper HW, Phinney KW, Thienpont LM, Coates PM, Vitamin D Standardization Program VDSP. Vitamin D status as an international issue: national surveys and the problem of standardization. *Scand J Clin Lab Invest* 2012;243:32–40.
3. Tai SS-C, Bedner M, Phinney KW. Development of a candidate reference measurement procedure for the determination of 25-hydroxyvitamin D3 and 25-hydroxyvitamin D2 in human serum using isotope-dilution liquid chromatography-tandem mass spectrometry. *Anal Chem* 2010; 82:1942–8.
4. Stepman HCM, Vanderroost A, Van Uytanghe K, Thienpont LM. Candidate reference measurement procedures for serum 25-Hydroxyvitamin D3 and 25-Hydroxyvitamin D2 by using isotope-dilution liquid chromatography-tandem mass spectrometry. *Clin Chem* 2011;57:441–8.
5. Mineva EM, Schleicher RL, Chaudhary-Webb M, Maw KL, Botelho JC, Vesper HW, et al. A candidate reference measurement procedure for quantifying serum concentrations of 25-hydroxyvitamin D3 and 25-hydroxyvitamin D2 using isotope-dilution liquid chromatography-tandem mass spectrometry. *Anal Bioanal Chem* 2017;407:5615–24.
6. Stöckl D, Sluss PM, Thienpont LM. Specifications for trueness and precision of a reference measurement system for serum/plasma 25-hydroxyvitamin D analysis. *Clin Chim Acta* 2009;408:8–13.
7. Fraser CG. The 1999 Stockholm consensus conference on quality specifications in laboratory medicine. *Clin Chem Lab Med* 2015;53: 837–40.
8. Sempos CT, Betz JM, Camara JE, Carter GD, Cavalier E, Clarke MW, et al. General steps to standardize the laboratory measurement of serum total 25-hydroxyvitamin D. *JAOAC Int* 2017;100:1230–3.
9. CDC. CDC Vitamin D Standardization-Certification Program (CDC VDSCP) Certified total 25-hydroxyvitamin D procedures.
10. Panteghini M, Braga F, Camara JE, Delatour V, Van Uytanghe K, Vesper HW, et al. Optimizing available tools for achieving result standardization: value added by joint committee on traceability in laboratory medicine (JCTLM). *Clin Chem* 2021;67:1590–605.

11. Sandberg S, Fraser CG, Horvath AR, Jansen R, Jones G, Oosterhuis W, et al. Defining analytical performance specifications: consensus statement from the 1st strategic conference of the European Federation of clinical chemistry and laboratory medicine. *Clin Chem Lab Med* 2015;53:833–5.
12. Carobene A, Strollo M, Jonker N, Barla G, Bartlett WA, Sandberg S, et al. Collections from healthy volunteers for biological variation estimates' update: a new project undertaken by the working group on biological variation established by the European Federation of clinical chemistry and laboratory medicine. *Clin Chem Lab Med* 2016;54:1599–608.
13. Cavalier E, Fraser CG, Bhattoa HP, Heijboer AC, Makris K, Ulmer CZ, et al. Analytical performance specifications for 25-hydroxyvitamin D examinations. *Nutrients* 2021;28:13.
14. Fraser CG. Reference change values. *Clin Chem Lab Med* 2012;50:807–12.
15. Fabregat-Cabello N, Farre-Segura J, Huyghebaert L, Peeters S, Le Goff C, Souberbielle JC, et al. A fast and simple method for simultaneous measurements of 25(OH)D, 24,25(OH)2D and the vitamin D metabolite ratio (VMR) in serum samples by LC-MS/MS. *Clin Chim Acta* 2017;473:116–23.
16. Elsenberg EHAM, ten Boekel E, Huijgen H, Heijboer AC. Standardization of automated 25-hydroxyvitamin D assays: how successful is it? *Clin Biochem* 2017;50:1126–30.
17. Wise SA, Camara JE, Semplos CT, Lukas P, Le Goff C, Peeters S, et al. Vitamin D standardization program (VDSP) intralaboratory study for the assessment of 25-hydroxyvitamin D assay variability and bias. *J Steroid Biochem Mol Biol* 2021;212:105917.
18. Arnaud J, Constans J. Affinity differences for vitamin D metabolites associated with the genetic isoforms of the human serum carrier protein (DBP). *Hum Genet* 1993;92:183–8.
19. Heijboer AC, Blankenstein Ma., Kema IP, Buijs MM. Accuracy of 6 routine 25-hydroxyvitamin D assays: influence of vitamin D binding protein concentration. *Clin Chem* 2012;58:543–8.
20. Cavalier E, Lukas P, Bekaert A-C, Peeters S, Le Goff C, Yayo E, et al. Analytical and clinical evaluation of the new fujirebio lumipulse non-competitive assay for 25(OH)-Vitamin D and three immunoassays for 25(OH)D in healthy subjects, osteoporotic patients, third trimester pregnant women, healthy African subjects. *Clin Chem Lab Med* 2016;54:1347–55.
21. Cavalier E, Lukas P, Crine Y, Peeters S, Carlisi A, Le Goff C, et al. Evaluation of automated immunoassays for 25(OH)-vitamin D determination in different critical populations before and after standardization of the assays. *Clin Chim Acta* 2014;431:60–5.
22. Makris K, Bhattoa HP, Cavalier E, Phinney K, Semplos CT, Ulmer CZ, et al. Recommendations on the measurement and the clinical use of vitamin D metabolites and vitamin D binding protein - a position paper from the IFCC committee on bone metabolism. *Clin Chim Acta* 2021;517:171–97.

Supplementary Material: This article contains supplementary material (<https://doi.org/10.1515/cclm-2025-0554>).